

Certificate Of Processing

Prepared for **QUANTUM TATTOO INK LLC**



Gamma Process Run ID **117369A**

<u>Product Code</u>	<u>Product Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
Tattoo Ink	0001	87	CS

Processing Run Start Date/Time: 30-Aug-2019 08:32:29 am Approx. Downtime (hours): 0.62
Processing Run End Date/Time: 30-Aug-2019 01:08:33 pm

Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	30.4
Maximum Specified Dose (kGy):	50.0	Maximum Delivered Dose (kGy):	47.2

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest



Reviewed and E-Signed By
Uprety, Amit (QS & RC Analyst I)

Document Content Revision: 1

Signed On 9/3/2019 at 12:14 PM
UTC / GMT Offset (hh:mm): -7:00

Processing Location:

STERIS
1000 South Sarah Place
Ontario, CA 91761
Phone: 909-390-9942
Fax: 909-390-7854

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.