

STERIS: Gamma Certificate Of Processing

Prepared for: QUANTUM TATTOO INK LLC (9098)

Gamma Process Run ID 1113-12984A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
Quantum Tattoo Ink LLC	0008	71	Case

Processing Run Start Date: 06-Jan-2023 5:17 PM

Approx. Downtime (hours): 0.98

Processing Run End Date: 07-Jan-2023 12:33 AM

Minimum Specified Dose (kGy): 25.0

Minimum Delivered Dose (kGy): 30.2

Maximum Specified Dose (kGy): 50.0

Maximum Delivered Dose (kGy): 43.0

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information

0008

Signature Manifest

Reviewed and E-Signed By: **Jessica Bean (Quality Analyst)**

Date/Time E-Signed: 08-Jan-2023 3:25 PM

Document Content Revision: 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

1000 Sarah Place
Ontario CA 91761
United States