

STERIS: Gamma Certificate Of Processing

Prepared for: QUANTUM TATTOO INK LLC (9098)

Gamma Process Run ID 1113-10405A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
Quantum Tattoo Ink LLC	0007	131	Case

Processing Run Start Date: 25-Jul-2022 11:16 AM

Approx. Downtime (hours): 1.67

Processing Run End Date: 25-Jul-2022 8:24 PM

Minimum Specified Dose (kGy): 25.0

Minimum Delivered Dose (kGy): 30.9

Maximum Specified Dose (kGy): 50.0

Maximum Delivered Dose (kGy): 45.6

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information

0007

Signature Manifest

Reviewed and E-Signed By: **Renee Rodriguez (QS/RC Technician)**

Date/Time E-Signed: 28-Jul-2022 4:46 AM

Document Content Revision: 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

1000 Sarah Place
Ontario CA 91761
United States